

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130135

Supervisor/Operator Performing the Verification Procedure:

Name: Trp. J. Santora

ID# 2276

Date: 08/6/2011

A

Agency: Alaska State Troopers

Phone # 656-1233

Instrument Location: Galena Police Department

B

Alco S/N: 78855

Target Value: .081

High Pressure: 800

Alco Test Values .080

.080

1st Alco

2nd Alco

Signature

(OVER)

(Do Not write in the area below)

CRF 8/22/11

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.

(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz


Scientific Director

State Breath Alcohol Program

Subscribed and sworn before me this 2 day of Sept, 2011. (Signature)

BT9

Nikki Roth
Nikki Roth, Notary Public
My Commission Expires With Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130135

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130135

AUGUST 06, 2011

OPERATOR'S NAME:
SANTORA JEANNINE L
OPERATOR'S NUMBER: 2276
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :

OF CAL
O.L. #: 0000000
DEPT/AGENCY: GALD
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .001
ALCO S/N: 78855

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.93 in		
ALCO TARGET	.001	16:52
BLANK TEST	.000	16:53
INTERNAL STANDARD	VERIFIED	16:53
ALCO TV 29.93 in	.000	16:54
BLANK TEST	.000	16:55
SUBJECT SAMPLE	.000	16:55
BLANK TEST	.000	16:56
ALCO TV 29.93 in	.000	16:56
BLANK TEST	.000	16:57

CRF 8/22/11

CRF 8/22/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130135

AUGUST 06, 2011
TIME 16:59

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.93 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~@